

HOEK TRIAXIAL CELL

Model HTC

APPLICATIONS

The Hoek Triaxial Cell is designed to determine the tri-axial strength of diamond drill cores of rock or concrete. Tests carried out on a series of samples under different confining pressures allow the user to determine:

- Strength and elastic properties
- Shear strength at different confining pressures
- Angle of shearing resistance and cohesion
- Modulus of elasticity and Poisson's ratio

The knowledge of these parameters is fundamental for the design of all important underground works such as powerhouses, storage facilities and in the quality assessment of existing concrete dams and structures.

Special Hoek Triaxial Cells can also be used instead of the biaxial chamber in the determination of the in-situ rock stress with the borehole deformation gauge.

DESCRIPTION

The HTC triaxial cell consists of a hollow steel cylinder with threaded removable ends. A urethane rubber sleeve incorporating U-shaped seals to form a pressurization chamber for the hydraulic fluid is mounted within the cell. Inside the cell is installed an inlet fitted with a 9.525 mm (3/8") quick-connect and an outlet with a plug for the saturation of the pressurization chamber.

At each end of the cell is a spherical seat, which will apply an axial load to the flattened ends of the sample. This pair of seats is made of tool steel hardened and ground to ASTM (D2664) standard. The clearance gap between the seats and the removable ends of the cell is sufficient to allow the passage of strain gauge leads.

TEST PROCEDURE

First, the ends of the rock specimen, with a minimum length to diameter ratio of 2, must be ground flat within a parallelism of 0.025 mm to 0.012 mm, depending on the diameter of the sample.



After saturation of the pressurization chamber, the cylindrical rock sample is inserted into the chamber within the confining membrane. The two spherical seats are positioned so that the rock core lies centrally in the triaxial chamber.

After applying a small confining pressure to hold the rock core in place, the cell with its spherical seats is placed in a loading frame and a small axial load is applied to hold the system firmly in place.

To determine Poisson's ratio, two orthogonal strain gauges are glued on the rock core. The strain gauges are read during the test with a Wheatstone bridge readout. The triaxial test may then be run after adjustment of the confining pressure to the required value.

FEATURES

- Easy to operate
- Models available for eight different core diameters up to 152 mm (6 inches)
- High pressure capability
- ASTM-compliant

INTERPRETATION

With the use of strain gauges, triaxial tests run on the same type of rocks at different confining pressures allow the determination of the failure envelope with the angle of friction and cohesion, the modulus of elasticity and Poisson's ratio.

SPECIFICATIONS

DESCRIPTION	SIZE ¹					
	EWG	AWG	BWG	NQ	NWG	HQ
Triaxial cell complete with one quick connect and one plug	0.845 in. (21.5 mm)	1.185 in. (30.1 mm)	1.655 in. (42.0 mm)	1.875 in. (47.6 mm)	2.155 in. (54.7 mm)	2.500 in. (63.5 mm)
SPECIAL HOEK CELLS						
Special Hoek triaxial cell for 144.7 mm (5.7 in.) core without platens, to use for overcore testing			Loading length: 400 mm (15.7 in.) Max. pressure: 70 MPa (10 000 psi)			
144.7 mm (5.7 in.) diameter membrane						
Special Hoek triaxial cell for 152.4 mm (6.0 in.) core with flat platens			Loading length: 380 mm (15.0 in.)			
152.4 mm (6.0 in.) diameter membrane						

¹ Other sizes available upon request.

ACCESSORIES ²

- Spherical seat platen
- Membrane
- Set of two complete permeability end caps with distance block (for L/D = 1)
- Load spreader (set of 2)
- Flat platen
- Support collar
- Straight pore pressure platen
- Hydraulic pump 0–10 000 psi with pressure gauge 0–10 000 psi, hose and male quick connect to mate with cell

² None of these, including the membrane, come with the cell; please mention desired accessories with ordering.